

QIAGEN Validation Report

Developmental validation of the Investigator[®] ESSplex SE Plus Kit

The QIAGEN[®] Investigator ESSplex SE Plus Kit is intended for molecular biology applications in forensic, human identity, and paternity testing. It is used for multiplex PCR in forensic casework. It was developed specifically for rapid and reliable generation of DNA profiles from blood, buccal swabs, and forensic stains.

The performance of the Investigator ESSplex SE Plus Kit was evaluated with regard to various sample types and conditions commonly encountered in forensic and parentage laboratories.

The validation study was based on the recommendations of the European Network of Forensic Science Institutes (ENFSI) (1) and the Revised Validation Guidelines of the Scientific Working Group on DNA Analysis Methods (SWGDM) (2). The assay makes use of well-established methodologies for forensic DNA analysis. It co-amplifies the 15 polymorphic STR markers recommended by the European Network of Forensic Science Institutes (ENFSI) and the European DNA Profiling Group (EDNAP) as the new European Standard Set of loci (D1S1656, D2S441, D2S1338, D3S1358, D8S1179, D10S1248, D12S391, D16S539, D18S51, D19S433, D21S11, D22S1045, FGA [FIBRA], TH01 [TC11], and vWA), plus SE33 [ACTBP2], and the gender-specific Amelogenin (3). These genetic loci have been characterized in numerous studies by other laboratories (4–6). Optimal reaction conditions were established, and the effects of variations in those conditions were assessed. The kit was tested in house and at independent external forensic laboratories.

The effect of variations in buffer component concentrations is described using the example of the MgCl₂ concentration (page 2). The robustness of the assay regarding variations in PCR cycling conditions and on a range of different thermal cyclers and genetic analyzers was investigated (pages 4–9). The effects of using increased or decreased PCR cycle numbers were analyzed (page 5). Sensitivity was addressed by amplifying DNA of known concentration in a range typically encountered in forensic casework analysis (page 10).



Specific issues that can arise during forensic casework were investigated, such as the ability to obtain results from samples that have been subjected to adverse environmental conditions. This was tested using DNA spiked with several types of inhibitors or degraded DNA (pages 12 and 16). Cross-reactivity with non-human DNA (page 17) was also assessed. The reproducibility of the results was verified (page 25).

Results of developmental validation

The validation study has been performed at the QIAGEN R&D department. All of the electropherograms shown were generated on an Applied Biosystems® 3500™ Genetic Analyzer. The standard conditions specified in the Investigator ESSplex SE Plus Handbook were used for the electrophoresis. Unless stated otherwise, a GeneAmp® PCR System 9700 with Gold-plated Silver 96-Well Block was used for amplification. Data were analyzed using QIAGEN Investigator IDproof Software, v2.0 or Applied Biosystems GeneMapper® ID-X software, v1.2.

Reaction conditions

Reaction conditions were established for an optimal performance in terms of sensitivity, specificity, and reproducibility. This required the optimization of all critical buffer components. The final composition of the Fast Reaction Mix is a robust buffer system that tolerates differences in the concentration of individual buffer components without a decrease in overall amplification performance.

For example, various concentrations of $MgCl_2$, one of the critical buffer components, were added to a standard reaction with 500 pg Control DNA XY13 (Figure 1). The assay yielded robust results within an $MgCl_2$ concentration range of $\pm 30\%$ of the optimum.

Under these conditions, amplification of the markers was well balanced and no dropout or nonspecific amplification occurred. Similar experiments were performed to evaluate optimal concentration for other buffer components, such as the hot-start DNA polymerase, dNTPs, and BSA (data not shown).




Figure 1. Influence of MgCl₂ concentration. 500 pg Control DNA XY13 was amplified on a GeneAmp PCR System 9700 under standard conditions (30 cycles). Each bar represents an average peak height of duplicates. Bar colors represent the fluorescent dye labels of individual markers, grey bars show the average of all markers.

Effect of PCR annealing temperature variations

Specificity, sensitivity, and robustness are critical in forensic casework, and all are affected by the annealing temperature (T_m). Since the actual T_m may vary depending on cyclor conditions, the assay was validated in a range surrounding the optimal T_m of the Investigator ESSplex SE Plus reaction (61°C).

Annealing temperatures between 56°C and 66°C were applied to the amplification of 500 pg Control DNA XY13 and PCR was performed on an Eppendorf® Mastercycler® ep instrument. Reactions using annealing temperatures between 56°C and 64.4°C resulted in full profiles. Good interlocus balance was observed for the temperature range of 59.2–62.8°C. However, the average peak height of markers was best for the conditions closest to the actual annealing temperature of 61°C (Figure 2). First dropouts were observed at 65.7°C. No nonspecific PCR products were observed above an annealing temperature of 57.5°C. In order to ensure optimal performance of the assay, we strongly recommend regular calibration of thermal cyclers.

Similar sets of experiments were performed to set optimal temperatures and hold times for all PCR protocol steps (data not shown).




Figure 2. Effect of variations in the PCR annealing temperature. 500 pg Control DNA XY13 was amplified on a GeneAmp PCR System 9700 under standard conditions (30 cycles). Each bar represents an average peak height of duplicates. Bar colors represent the fluorescent dye labels of individual markers, grey bars show the average of all markers.

Effect of different cycle numbers

Altering the number of PCR cycles can be used to adapt the reaction conditions to varying DNA template concentrations. Cycle numbers can be either increased to enhance amplification signals when working with low-copy-number DNA, or decreased to speed up the protocol when the DNA sample is abundant (e.g., for database samples).

Cycle numbers were increased to 32 or 34 for reactions containing 32, 16, or 8 pg of Control DNA XY5, and the numbers of called alleles, as well as peak heights, were compared to a standard 30-cycle protocol (Figure 3). As expected, signal intensities of amplified products increase with higher cycle numbers. However, it should be noted that an increase in cycle number to more than 30 will not necessarily result in more information obtained from the low-template-DNA sample. Furthermore, because of stochastic effects, increased peak imbalances or dropouts may in general be observed for low-copy-number samples (with 100 pg or less of template DNA), regardless of any increase in cycle numbers. In this experiment, using a threshold of 50 RFU for allele calling, the number of allelic dropouts due to stochastic effects was not significantly reduced when more PCR cycles were applied (Figure 3). Note that the sensitivity of the capillary electrophoresis instrument and the setting of the threshold of detection strongly influence the outcome of an assessment of the use of increased cycle numbers. Therefore, such protocol adaptations must be evaluated by individual laboratories using their instrumentation and analysis. Adapting the injection time of the capillary electrophoresis instrument or using post-PCR purification (e.g., the QIAGEN MinElute[®] PCR Purification Kit) provide alternative methods for increasing sensitivity.




Figure 3. Effect of different cycle numbers on mean peak height and number of called alleles. DNA XY5 was used as template using different amounts, as indicated. Samples were run in duplicate and numbers of detected PCR products (indicated above the bars; 64 PCR products expected) and their peak heights calculated. 50 RFU was used as a threshold for detection.

In particular, for reference samples such as buccal swabs, where DNA can be extracted in abundance, reduced cycle numbers may be used to streamline the laboratory workflow. Here, cycle numbers were decreased to 24, 26, or 28 for reactions containing 2.5 ng or 10 ng of template DNA (Figure 4). All reactions resulted in robust amplification and full profiles were obtained using a threshold of 50 RFU for allele calling. However, 28 cycle amplifications of 10 ng DNA gave rise to pull-up peaks when applying samples to analysis without prior dilution of the PCR product. An example electropherogram for a 24 cycle amplification of 10 ng template DNA is shown in Figure 5, next page.




Figure 4. Effect of decreased cycle numbers on mean peak heights. 2.5 ng or 10 ng of Control DNA XY5 were subjected to amplification using 24, 26, or 28 PCR cycles. Samples were run in duplicate and average peak heights calculated.




Figure 5. Reduction of PCR cycle numbers with elevated template DNA amounts. Results for amplification of 10 ng Control DNA XY5 using 24 instead of standard 30 PCR cycles. Data were analyzed using QIAGEN Investigator IDproof Software.

Effect of different cycler types

Several PCR thermal cyclers were tested with the Investigator ESSplex SE Plus Kit to demonstrate kit robustness independent of the instrument (Figure 6, next page). 2 ng to 32 pg Control DNA XY13 was used as a PCR template. The reaction took place under standard conditions (30 cycles) and was performed with the following thermal cyclers.

- GeneAmp PCR System 9700 with Aluminum 96-Well Block (Applied Biosystems Inc., Foster City, CA, USA)
- GeneAmp PCR System 9700 with Silver or Gold-plated Silver 96-Well Block (Applied Biosystems Inc., Foster City, CA, USA)
- MJ Research DNA Engine® PTC-200 Peltier Thermal Cycler (Bio-Rad Laboratories GmbH, Munich, Germany)
- UNO-Thermo Block (Biometra Biometra biomedizinische Analytik GmbH, Göttingen, Germany)
- QIAGEN Rotor-Gene® Q




Figure 6. Performance on different PCR thermal cyclers. DNA XY13 was used as template using different amounts, as indicated. Each sample was run in duplicate. A standard 30-cycle protocol was used. Average peak heights across all markers are shown.

Comparable mean peak heights were obtained for all of the tested PCR cyclers. No significant differences in intra- and inter-locus balance or nonspecific amplification was observed on any of the thermal cyclers. In the case of the GeneAmp PCR System 9700, it is essential to use the “Standard mode” in order to achieve optimal results. Do not use the “9600 emulation mode”.

Effect of different genetic analyzers

Several genetic analyzers were tested with the Investigator ESSplex SE Plus Kit in order to demonstrate its robustness. 500 pg Control DNA XY13 and Control DNA 9948 were each amplified in duplicate. The reaction took place under standard conditions and the PCR products were tested on all genetic analyzers in parallel:

- Applied Biosystems 3500 Genetic Analyzer
- Applied Biosystems 3130 Genetic Analyzer
- ABI PRISM® 310 Genetic Analyzer

The electropherograms in Figure 7 show comparable mean peak heights for all of the tested genetic analyzers. No imbalance, dropouts, or preferential amplification for the STR systems was observed on either genetic analyzer.

Similar overall peak heights and a comparable balance of peaks between individual markers of the same color channel, as well as between different channels, were observed (Figure 7). Variations reflected differences in sensitivity between individual instruments, e.g., due to laser power, array life, and matrix calibration effects. Limitations of genetic analyzer sensitivity must be addressed by individual laboratories during an internal validation study using the instrumentation in their laboratory.




Figure 7. Analysis of PCR products on different genetic analyzers. 500 pg Control DNA 9948 and XY13 was amplified on a GeneAmp PCR System 9700 in duplicate. Samples were run according to the *Investigator ESSplex SE Plus Handbook*. Average peak heights for duplicates of both DNAs are shown.

Sensitivity

The Investigator ESSplex SE Plus Kit is designed to work robustly over a range of DNA quantities. The recommended amount of input DNA to yield good quality STR profiles is 500 pg, based on real-time PCR quantification of human DNA, e.g., using the QIAGEN Investigator Quantiplex Kit. In particular, for heavily degraded DNA, the use of increased template amounts may improve results (see page 16).

Control DNA XY13 was serially diluted from 2 ng to 8 pg per reaction. Full profiles (34 PCR products) were consistently obtained at 125 pg using the standard conditions specified in the *Investigator ESSplex SE Plus Handbook*. Occasional allele dropouts were found due to stochastic effects when ≤ 63 pg DNA was used as template. As expected, the number of dropouts increases with decreasing DNA concentration. See “Effect of different cycle numbers” and Figure 3 for expected numbers of detected alleles with low-copy-number samples.

Issues with very high amounts of DNA

Although DNA template amounts >1 ng do not cause issues during PCR amplification, those samples still can result in inaccurate or unusable data. Depending on the instrumentation and settings used for capillary electrophoresis, fluorescence intensity may go off the scale.

“Off-scale” peaks are often accompanied by “pull-up” peaks. Furthermore, higher stutter-peak height ratios may be observed as the signal from the main peak gets saturated. Finally, “split peaks” may occur as a result of incomplete +A nucleotide addition.

Poor STR profiles resulting from high DNA concentration can be improved by reamplifying a sample using less template DNA.

Issues with very low amounts of DNA

Amplification of less than ~ 100 pg DNA may lead to incomplete profiles, lacking one or more alleles. Furthermore, low allele copy numbers in the PCR can result in an unbalanced amplification of the alleles due to stochastic fluctuation.

These effects can be handled by either performing multiple amplifications of the same sample to create a consensus profile, or by adding the maximum volume of the DNA template to the PCR in order to get the best possible result from a single reaction. For correct interpretation of samples having only small amounts of DNA, it is crucial to minimize background fluorescence, e.g., caused by dye artifacts. Figure 8 (next page) shows an example of a no-template amplification. Peak heights of amplification products and the level of background noise depend on the instruments and settings of individual laboratories. We therefore recommend evaluating a suitable threshold for detection of alleles based on results obtained during an internal validation in the laboratory.




Figure 8. Sensitivity study. Serial dilutions of Control DNA XY13 were analyzed. The amounts of DNA indicated were used as template for amplification. **Note:** Y-axis scales were adjusted for best fit. The lowest panel shows a no-template (negative) control to illustrate the expected level of background fluorescence.

Performance with simulated inhibition

If the DNA extraction from forensic casework samples is done using inappropriate methods, Investigator ESSplex SE Plus assay performance may be compromised, although the reaction chemistry has been optimized for increased inhibitor tolerance. QIAGEN sample preparation technology is recommended for extraction, as it yields pure DNA free from inhibitors.

To test the robustness of the kit in the presence of inhibitors, assays were run in the presence of the following inhibitors, chosen to mimic challenging forensic sample types:

- Humic acid: A principal component of humic substances and has an inhibitory effect on PCR. It may be co-extracted from forensic samples collected from soil
- Hematin: Formed by the oxidation of heme, the main component of blood. It has been identified as a PCR inhibitor in DNA samples extracted from bloodstains
- Tannic acid: Typically present in leather. During extraction of casework evidence, tannic acid may get co-extracted with DNA
- Indigo carmine: The colour of blue denim (e.g., jeans)
- Collagen: The main protein compound of many tissues
- Calcium: Released during lysis of bones

Results are shown in Figures 9–12. Figure 13 shows an overview of ESSplex SE Plus inhibitor resistance.




Figure 9. Electropherogram of 500 pg Control DNA XY13 amplified in the presence of 150 ng/ μ l humic acid. Data were analyzed using QIAGEN Investigator IDproof Software.




Figure 10. Electropherogram of 500 pg Control DNA XY13 amplified in the presence of 1000 μ M hematin. Data were analyzed using QIAGEN Investigator IDproof Software.




Figure 11. Electropherogram of 500 pg Control DNA XY13 amplified in the presence of 8 mM indigo carmine. Data were analyzed using QIAGEN Investigator IDproof Software.




Figure 12. Electropherogram of 500 pg Control DNA XY13 amplified in the presence of 150 ng/μl collagen. Data were analyzed using QIAGEN Investigator IDproof Software.




Figure 14. Effect of high levels of tannic acid. 500 pg Control DNA XY13 was amplified in the presence of 500 ng or 1000 ng/ μ l tannic acid. Arrows indicate PCR products showing split peaks caused by partial adenylation.

Stability with degraded DNA

Casework evidence has often been exposed to adverse environmental conditions, e.g., UV light, humidity, or microbial growth. Human DNA extracted from those samples therefore may show varying degrees of degradation.

Artificially degraded male genomic DNA was used for validation. DNA was sheared using adaptive focused acoustics (Covaris system, Covaris Inc., Woburn, MA, USA) to average fragment lengths of 500 bp, 300 bp, or 150 bp according to the manufacturer's instructions. Fragment lengths were verified on an agarose gel and 500 pg DNA was used as template for amplification (all samples in duplicate) (Figure 15A). Full profiles were obtained with DNA of 300 bp and 500 bp average length. As expected, loci with larger amplicon sizes become affected as DNA degradation progresses. At an average fragment length of 150 bp, 104 out of 132 expected peaks (79%) were detected using a threshold of 50 RFU (Figure 15B). Increased amounts of template (up to 2 ng), can be used to improve results for heavily degraded DNA.




Figure 15. Degraded DNA. Male DNA was sheared to the defined average fragment length indicated. **A** Agarose gel analysis of degraded DNA. **M**: DNA size standard markers, **C**: Untreated DNA. 500, 300, 150: average fragment lengths in base pairs. **B** STR analysis of untreated control and degraded DNA samples. 500 pg template DNA was used. Full profiles were obtained down to 300 bp average fragment length.

Species specificity

Non-human DNA can be present in forensic casework samples. It is critical that assays show no cross-reactivity between species. To verify Investigator ESSplex SE Plus species specificity for human DNA, DNA from other species was each tested following the standard assay protocol (Figure 16, next page).

Besides common pets and farm animals, some primates were also tested. As expected for primates, amplification of some products is possible. Chimpanzee, bonobo, orangutan, and gorilla give rise to several peaks within marker ranges in all channels, some of which match the size of human STR products. Macaque DNA produced an Amelogenin X-peak and further off-ladder peaks in the FAM™, BTG, and BTR panels.

Dog, rat, pig, and goat showed each 2 or fewer low-level, off-ladder peaks in the FAM and BTG panels (<100 RFU). The highest off-ladder peak of 1600 RFU was produced with 2.5 ng sheep DNA in the BTG channel at 429 bp.

Allele calls with peak heights between 50 and 130 RFU were observed for rat (D10S1248, allele 13) and horse (D16S539, allele 13; D1S1656, allele 13; D2S441, allele 14).




Figure 16. Representative results of the species specificity assessment. 500 pg primate DNA and 2.5 ng DNA from all other species was used as template.

Mixture studies

Evidence samples are frequently composed of more than one individual's DNA. For correct interpretation of results from mixtures, it is important to know the limit of the minor contributing component that still can be resolved.


Samples were created by mixing control DNA XY5 and XY13 in ratios of 1:1, 3:1, 7:1, 10:1, 15:1, and vice versa (see Table 1 for genotypes of mixed samples). The total amount of mixed DNA used in this study was 500 pg; a 15:1 mixture thus contains 31 pg of the minor component DNA and 469 pg of the major component (Table 2). The limit of detection of the minor component was determined by analyzing non-overlapping alleles of both DNAs. All expected alleles were found for minor components of 3:1 and 7:1 mixtures. 10:1 and 15:1 typically resulted in partial profiles of the minor component. An example for 1:10 mixtures is shown in Figure 17 (pages 21–24). Since these contain ≤ 50 pg of the minor component, the results are in concordance with the sensitivity for single-source samples reported here. In order to increase the sensitivity for the minor component, higher overall DNA amounts may be used if the amount of available DNA is not limited. See "Sensitivity", page 10, for general considerations.

Table 1. Genotypes of DNAs used for mixed samples


Locus	XY5 genotype	XY13 genotype
Amelogenin	X/Y	X/Y
D1S1656	16/16	16/17.3
D2S441	10/14	10/11
D2S1338	16/19	17/23
D3S1358	14/16	15/16
D8S1179	12/15	13/14
D10S1248	13/15	14/15
D12S391	17/19	16/18
D16S539	9/12	11/12
D18S51	13/15	14/15
D19S433	13/14	13/14
D21S11	28/31	28.2/33.2
D22S1045	15/16	17/18
FGA	20/25	22/23
SE33	15/21.2	14/24.2
TH01	6/7	6/7
vWA	16/16	15/16

Table 2. Amounts of DNA template in mixtures


Mixture ratio	Major component	Minor component
1:1	250 pg	250 pg
3:1	375 pg	125 pg
7:1	437 pg	63 pg
10:1	450 pg	50 pg
15:1	469 pg	31 pg

A


Expected bp positions of uncovered minor component peaks are indicated by red arrows. Amelogenin, TH01: Heterozygote + heterozygote, 4 overlapping alleles. Major component completely covers minor component. D3S1358: Heterozygote + heterozygote, 2 overlapping alleles, 2 alleles overlapping with stutter positions. Minor component is covered by major component, or visible as elevated stutter. vWA: Homozygote + heterozygote, 2 overlapping alleles, 1 allele overlapping with stutter position. Minor component is covered by major component, or visible as elevated stutter. D21S11: Heterozygote + heterozygote, no overlapping alleles. All minor component peaks can be assigned.

B

Expected bp positions of uncovered minor component peaks are indicated by red arrows. D16S539: Heterozygote + heterozygote, 2 overlapping alleles, 1 allele overlapping with stutter position. The non-overlapping minor component peak can be assigned. D1S1656: Homozygote + heterozygote, 2 overlapping alleles. The non-overlapping minor component peak can be assigned. D19S433: Heterozygote + heterozygote, 4 overlapping alleles. Major component completely covers minor component. SE33: Heterozygote + heterozygote, one allele overlapping with stutter position. Minor component is covered by major component, or visible as elevated stutter. The 3 non-overlapping minor component peaks can be assigned.

C

Expected bp positions of uncovered minor component peaks are indicated by red arrows. D10S1248: Heterozygote + heterozygote, 2 overlapping alleles, 2 alleles overlapping with stutter position. Minor component is covered by major component, or visible as elevated stutter. D22S1045: Heterozygote + heterozygote, 1 allele overlapping with stutter position. Minor component is covered by major component, or visible as elevated stutter. The 3 non-overlapping minor component peaks can be assigned, but thereof one peak height is very close to 50 RFU threshold. D12S391: Heterozygote + heterozygote, 3 alleles overlapping with stutter positions. Minor component is covered by major component, or visible as elevated stutter. The non-overlapping minor component peak can be assigned. D8S1179: Heterozygote + heterozygote, 2 alleles overlapping with stutter positions. Minor component is covered by major component, or visible as elevated stutter. 2 non-overlapping minor component peaks can be assigned. D2S1338: Heterozygote + heterozygote, one allele overlapping with stutter position. Minor component is covered by major component, or visible as elevated stutter. The 3 non-overlapping minor component peaks can be assigned.

D

Expected bp positions of uncovered minor component peaks are indicated by red arrows. D2S441: Heterozygote + heterozygote, 2 overlapping alleles. The 2 non-overlapping minor component peaks can be assigned. D18S51: Heterozygote + heterozygote, 2 overlapping alleles, 2 alleles overlapping with stutter positions. Minor component is covered by major component, or visible as elevated stutter. FGA: Heterozygote + heterozygote, no overlapping alleles. All 4 minor component peaks can be assigned.

Figure 17. Results of the mixture analysis. A–D shows 1:10 mixtures of DNA XY5 and XY13 (middle panels of each figure) and the corresponding single source samples as a reference (upper and lower panel of each figure). Red arrows indicate positions where a peak of the minor component DNA that does not overlap with a main peak, is to be expected. A brief description of the mixture situation given for each individual marker is given below the figures. All but one expected peaks were found with a threshold of detection of 50 RFU. In this example, only allele 18 of D22S1045 of DNA XY13 dropped out (Figure 17C, second panel).

Reproducibility: Concordance test

All primer binding sites of the Investigator ESSplex SE Plus Kit are the same as in the Investigator ESSplex SE assay. Minor changes (5' tailing of unlabeled primers) have been introduced to some sequences in order to optimize adenylation of PCR products and allow faster PCR cycling.

A concordance study with an internal DNA pool of 151 Caucasian samples in total was conducted. Identical DNA profiles were obtained for all samples using either the Investigator ESSplex SE, or the ESSplex SE Plus Kit, indicating that the minor primer changes between the two primer sets do not affect allele calls. The internal DNA pool had previously been analyzed using the Applied Biosystems AmpF ℓ STR $^{\text{®}}$ Identifiler $^{\text{™}}$ PCR Kit. Genotypes obtained for all 10 overlapping loci and Amelogenin were fully concordant for all but 3 samples. These nonconcordant samples were previously typed for D19S433 as homozygous allele 13, but revealed a heterozygous 12/13 genotype when analyzed with the ESSplex SE Plus Kit. No nonspecific amplification was observed in any of the samples.

Peak height ratios of heterozygous markers (Figure 18) and stutter peak heights (Figure 19, pages 26–28) were analyzed for 762 samples of the internal DNA pool.
















Figure 18. Box plot analysis of intra-locus peak height ratios. Values refer to the lower of the 2 peaks in a heterozygous sample. Boxes represent the middle 50% (interquartile range, [IQR]) of data, lines inside the boxes the median. Whiskers show data within 1.5 IQR, dots represent outliers.








Figure19. Stutter peak height ratios. Peak heights of forward stutters compared to main peaks were analyzed for 762 DNA samples from different donors.

Stutter peak heights are characteristic for each marker and the number of repeat motifs of an allele. TH01 in general shows the lowest stutter ratio of all STR markers of the Investigator ESSplex SE Plus Kit. Smaller alleles display lower stutter levels than longer alleles of the same marker. Only forward stutter positions were analyzed, as backward stutters usually are of very low height, although present.

References

QIAGEN maintains a large, up-to-date online database of scientific publications utilizing QIAGEN products. Comprehensive search options allow you to find the articles you need, either by a simple keyword search or by specifying the application, research area, title, etc.

For a complete list of references, visit the QIAGEN Reference Database online at www.qiagen.com/RefDB/search.asp or contact QIAGEN Technical Services or your local distributor.

Cited references

1. ENFSI Standing Committee for Quality and Competence (QCC). Validation and Implementation of (New) Methods. Ref. Code: QCC-VAL-001, Issue No. 001, 4 November 2006. http://www.enfsi.eu/get_doc.php?uid=144.
2. Revised Validation Guidelines of Scientific Working Group on DNA Analysis Methods (SWGDM) Forensic Science Communications, July 2004, Volume 6, Number 3. www.cstl.nist.gov/strbase/validation/SWGDAM_Validation.doc.
3. Gill, P., et al. (2006) The evolution of DNA databases-Recommendations for new European STR loci. *Forensic Sci. Int.* 156, 242.
4. Phillips, C., et al. (2011) Analysis of global variability in 15 established and 5 new European Standard Set (ESS) STRs using the CEPH human genome diversity panel. *Forensic Sci. Int. Genet.* 5, 155.
5. Butler, J.M., Shen, Y., and McCord, B.R., (2003) The development of reduced size STR amplicons as tools for analysis of degraded DNA. *J. Forensic Sci.* 48, 1054.
6. Lareu, M.V., et al. (1996) A highly variable STR at the D12S391 locus. *J. Legal Med.* 109, 134.

Ordering Information

Product	Contents	Cat. no.
Investigator ESSplex SE Plus Kit (100)	Primer mix, Fast Reaction Mix including HotStarTaq <i>Plus</i> DNA Polymerase, Control DNA, allelic ladder ESSplex SE Plus, DNA size standard (BTO), and RNase-free water	381545
Investigator ESSplex SE Plus Kit (400)	Primer mix, Fast Reaction Mix including HotStarTaq <i>Plus</i> DNA Polymerase, Control DNA, allelic ladder ESSplex SE Plus, DNA size standard (BTO), and RNase-free water	381547

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual. QIAGEN kit handbooks and user manuals are available at www.qiagen.com or can be requested from QIAGEN Technical Service or your local distributor.

Trademarks: QIAGEN®, Investigator®, HotStarTaq®, MinElute®, Rotor-Gene® (QIAGEN Group); 3500™, ABI PRISM®, AmpFISTR®, Applied Biosystems®, FAM™, GeneAmp®, GeneMapper®, Identifiler™ (Applied Biosystems); DNA Engine® (Bio-Rad Laboratories GmbH); Eppendorf®, Mastercycler® (Eppendorf AG). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.

© 2012 QIAGEN, all rights reserved.

www.qiagen.com

Australia = 1-800-243-800

Austria = 0800/281010

Belgium = 0800-79612

Canada = 800-572-9613

China = 021-51345678

Denmark = 80-885945

Finland = 0800-914416

France = 01-60-920-930

Germany = 02103-29-12000

Hong Kong = 800 933 965

Ireland = 1800 555 049

Italy = 800 787980

Japan = 03-5547-0811

Korea (South) = 1544 7145

Luxembourg = 8002 2076

The Netherlands = 0800 0229592

Norway = 800-18859

Singapore = 65-67775366

Spain = 91-630-7050

Sweden = 020-790282

Switzerland = 055-254-22-11

UK = 01293-422-911

USA = 800-426-8157

