

Product Sheet

QIAamp[®] DNA Mini QIAcube[®] Kit

For use on QIAcube Connect

QIAamp DNA Mini QIAcube Kit	(240)
Catalog no.	51326
QIAamp Mini Spin Columns	240
Rotor Adapters	240
Elution Tubes	240
Buffer AL*	2 x 33 ml
Buffer ATL	50 ml
Buffer AW1* (concentrate)	98 ml
Buffer AW2† (concentrate)	66 ml
Buffer AE	60 ml
Proteinase K	6 ml
Product sheet	1

* Contains chaotropic salt. Not compatible with disinfecting agents containing bleach.

† Contains sodium azide as a preservative.

Storage

QIAamp Mini Spin Columns and buffers can be stored dry at room temperature (15–25°C) for up to 1 year, if not otherwise stated on the label. Proteinase K solution can be stored at room temperature or at 2–8°C for up to 1 year without affecting performance.

Intended Use

The QIAamp DNA Mini QIAcube Kit is intended for molecular biology applications. This product is not intended for the diagnosis, prevention or

treatment of a disease. For more information about the automated procedure, see the relevant protocol sheet available at www.qiagen.com/MyQIAcube.

QIAcube Connect is designed to perform fully automated purification of nucleic acids and proteins in molecular biology applications. The system is intended for use by professional users trained in molecular biological techniques and the operation of QIAcube Connect.

Safety Information

When working with chemicals, always wear a suitable lab coat, disposable gloves and protective goggles. For more information, please consult the appropriate safety data sheets (SDSs). These are available online in convenient and compact PDF format at www.qiagen.com/Support/MSDS.aspx where you can find, view and print the SDS for each QIAGEN® kit and kit components.

<p>CAUTION</p> 	<p>CAUTION: DO NOT add bleach or acidic solutions directly to the sample-preparation waste.</p>
---	--

The sample-preparation waste contains guanidine hydrochloride from Buffers AL and AW1, which can form highly reactive compounds when combined with bleach. If liquid containing these buffers is spilt, clean with suitable laboratory detergent and water. If the spilt liquid contains potentially infectious agents, clean the affected area first with laboratory detergent and water, and then with 1% (v/v) sodium hypochlorite.

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual. QIAGEN kit handbooks and user manuals are available at www.qiagen.com or can be requested from QIAGEN Technical Services or your local distributor.

Trademarks: QIAGEN®, Sample to Insight®, QIAamp®, QIAcube® (QIAGEN Group). 1120411 HB-0394-003 © 2020 QIAGEN, all rights reserved.