

March 2016

## Important Note

Dear valued customer,

We would like to inform you about changes to the **Limit of Blank (LOB) and Limit of Detection (LOD) for the G12V mutation** for the **KRAS Pyro® Kit** (QIAGEN cat. no. 970460). These changes are based on several in-depth QIAGEN investigations.

The LOB has changed from 0.0% units to 4.9% units and the LOD has changed from 1.0% units to 8.0% units. The *KRAS Pyro Kit handbook* and the *KRAS Pyro Plug-in Quick-Start Guide* will be updated to reflect these changes. There are no changes to values for the other mutations.

If you have any questions, please contact your local QIAGEN Technical Services Department.

Best regards,

Your QIAGEN Team

Trademarks: QIAGEN®, Sample to Insight®, Pyro®, (QIAGEN Group).  
1101370 03/2016 HB-2092-001  
© 2016 QIAGEN, all rights reserved.