

Quick-Start Protocol

QIAprep& Buffer AB

The QIAprep& Buffer AB (cat. nos. 221515 and 221517) should be stored upon receipt at -30 to -15°C in a constant-temperature fridge and protected from light.

Further information

- *QIAprep& Viral RNA UM Kit Handbook*: www.qiagen.com/HB-2830
- Safety Data Sheets: www.qiagen.com/safety
- Technical assistance: support.qiagen.com

Notes before starting

- The QIAprep& Buffer AB is intended for increasing sensitivity of the QIAprep& Viral RNA workflow when using sample types such as neat/pure saliva and gargle samples or other sample types requiring pre-treatment.
- The QIAprep& Buffer AB is intended to be used in conjunction with a QIAprep& Viral RNA UM Kit (cat. nos. 221415 or 221417). Refer to these kits' handbook for safe and proper use.
- For samples such as nasal, nasopharyngeal, or oropharyngeal swabs that are stored in non-fixation transport media such as UTM, VTM, PBS, ESswabs®, Virocult™, or 0.9% NaCl, the use of the QIAprep& Buffer AB is not required; instead, refer to the *QIAprep& Viral RNA UM Kit Handbook*.
- The protocol in this Quick-Start Protocol includes a heat treatment step before the sample preparation step. QIAGEN cannot guarantee that this heat treatment step will inactivate 100% of viral particles. The inactivation of virus needs to be verified and validated by users.

Kit contents

QIAprep&amp; Buffer AB	4.1 ml	16.4 ml
Catalog no.	221515	221517
QIAprep& Buffer A	2x 1.3 ml	8x 1.3 ml
QIAprep& Buffer B	1.5 ml	4x 1.5 ml
Quick-Start Protocol	1	1

Procedure

Heat pre-treatment of samples with the QIAprep& Buffer AB

1. Before use, prepare the QIAprep& Buffer AB according to Table 1 and mix thoroughly.
2. Pre-dispense 6 µl of Buffer AB into a PCR tube or the well of a PCR plate
3. Add 18 µl of sample to the tube or well containing the Buffer AB. Mix by pipetting up and down at least twice.

Note: Saliva samples can have high viscosity. Heating the primary sample to 80°C for 10 min can lower viscosity of the sample and facilitate pipetting of saliva samples. Heating the primary sample to 80°C for 10 min does not replace the heating step at 95°C for 5 min in step 4.

4. Seal the tube/plate and incubate for 5 min at 95°C.
5. Briefly centrifuge the tubes. Gently mix by pipetting up and down at least twice, and transfer 8 µl of heat pre-treated sample into new PCR tube or well and proceed to step 6 below.

Table 1. Saliva Prep Buffer setup

Component	1 rxn	Final concentration
QIAprep& Buffer A	3.75 µl	1x
QIAprep& Buffer B*	2.25 µl	1x
Total reaction volume	6 µl	–

* contains Proteinase K

PCR Setup

6. Prepare a reaction mix according to Table 2 and mix thoroughly.
7. Add 12 µl of the reaction mix prepared in step 6 to 8 µl of sample prepared in step 5.
Mix by pipetting up and down at least twice. The complete reaction can be stored up to 1 h at room temperature or stored frozen at –30 to –15°C for the longer period.

Table 2. Reaction mix setup

Component	96/384-well block	Final concentration
Viral RNA Master Mix, 4x*	5 µl	1 x
20x primer–probe mix	1 µl	1 x
RNA IC Template + Assay, 10x*	2 µl	1 x
Human Sampling IC Assay, 20x*	1 µl	1 x
ROX Reference Dye (ABI instruments only)*	1 µl/0.1 µl†	1 x
Viral RNA UM Prep Buffer*	2 µl	1 x
RNase-Free Water*	Fill up to 12 µl	–
Prepared sample (after step 5)	8 µl	–
Total reaction volume	20 µl	–

* Kit component of QIAprep& Viral RNA UM Kit (cat. nos. 221414, 221415, and 221417). Refer to the *QIAprep& Viral RNA UM Handbook* for safe and proper use.

† To be used as a 20x concentrate for high-ROX dye cyclers (i.e., ABI PRISM® 7000, Applied Biosystems® 7300, 7900, and StepOne® Real-Time PCR Systems) and as a 200x concentrate for low ROX-dye cyclers (i.e., Applied Biosystems 7500, ViiA7™, and QuantStudio® Real-Time PCR Systems).

8. Important consideration:

Seal the plate/tube thoroughly to prevent cross-contamination. In case an adhesive film is used, make sure to apply pressure uniformly across the entire plate, to obtain a tight seal across individual wells.

Mix gently by vortexing for 10–30 s with medium pressure. Place the plate in different positions while vortexing, to ensure an equal contact with the vortex platform.

Centrifuge the plate/tube briefly to collect liquid at the bottom of the plate/tube.

9. Program the real-time cycler according to Table 3.

Note: Data acquisition should be performed during the annealing/extension step.

Table 3. Cycling conditions

Step	Time	Temperature	Ramp rate
RT-step	10 min	50°C	Maximal/fast mode
PCR initial heat activation	2 min	95°C	Maximal/fast mode
2-step cycling (40 cycles)			
Denaturation	5 s	95°C	Maximal/fast mode
Combined annealing/extension	30 s	58°C*	Maximal/fast mode

* Annealing temperatures can be adapted between 55–62°C depending on primer/probe set used. For further details on cycling conditions, primer/probe concentrations, and annealing temperature, visit the product page (www.qiagen.com/qiaprepandamp-resources).

10. Place the tubes or plates in the real-time cycler and start the cycling program.

11. For results interpretation, refer to the table “Possible outcome” in the *QIAprep& Viral RNA UM Kit Handbook*: www.qiagen.com/HB-2830.

Document Revision History

Date	Changes
04/2021	Initial release
05/2021	Revised the volume of both QIAprep& Buffers A and B in the Kit Contents table.



Scan QR code for the product page and supplementary protocols.

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual.

Trademarks: QIAGEN®, Sample to Insight® (QIAGEN Group); ESwab® (Copan Italia S.P.A.); Virocult™ (Medical Wire & Equipment Co.); ABI PRISM®, Applied Biosystems®, QuantStudio®, StepOne®, ViiA™ (Thermo Fisher Scientific or its subsidiaries). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.
1124638 05/2021 HB-2887-002 © 2021 QIAGEN, all rights reserved.