

Quick-Start Protocol

QIAseq® FastSelect™ –rRNA Plant with the NEBNext® Ultra II Directional Library Prep Kit

The QIAseq FastSelect Kits for –rRNA Plant (cat. nos. 334315, 334317, 334319) may be used with the NEBNext Ultra II Directional RNA Library Prep Kit for Illumina® (NEB cat. nos. E7760S and E7760L) to remove plant rRNA.

All components of QIAseq FastSelect should be stored in a constant-temperature freezer at –30 to –15°C. Under these conditions, the components are stable, without showing any reduction in performance and quality, until the date indicated on the box label.

Further information

- *QIAseq FastSelect –rRNA Plant Handbook*: www.qiagen.com/HB-2783
- Safety Data Sheets: www.qiagen.com/safety
- Technical assistance: support.qiagen.com

Notes before starting

- The NEBNext Ultra II Directional RNA Library Prep Kit for Illumina is required for use with this protocol.
- Refer to the *NEBNext Ultra II Directional RNA Library Prep Kit for Illumina Instruction Manual*.

Procedure

1. Vortex the tube(s) from the QIAseq FastSelect Kit, and then briefly centrifuge to collect residual liquid from the sides of the tubes.
2. Referring to Section 4 from the *NEBNext Ultra II Directional RNA Library Prep Kit for Illumina Instruction Manual*, perform the following in place of steps 4.1.1 through 4.1.4:
 - 2a. Assemble the fragmentation and priming reaction described in Table 1 on ice in a nuclease-free tube.

Table 1. NEBNext Ultra II Stranded fragmentation and priming mix

Component	Volume/reaction
Total RNA (5 ng – 1 µg)	4 µl
(lilac) NEBNext First Strand Synthesis Reaction Buffer*	4 µl
(lilac) Random primers*	1 µl
Total volume	9 µl

* From the NEBNext Ultra II Directional Library Prep Kit.

- 2b. To the assembled fragmentation and priming mix, add 1 µl of QIAseq FastSelect –rRNA Plant.
- 2c. Mix thoroughly by pipetting up and down several times, and then briefly centrifuge to collect residual liquid from the sides of the tubes.
- 2d. Incubate in a thermal cycler with a heated lid, as described in Table 2, according to your input RNA quality.

Important: Regardless of the time and temperature chosen in step 1, steps 2–9 must be performed.

Table 2. Combined NEBNext Ultra II fragmentation and FastSelect hybridization protocol

Step	Intact RNA (RIN >7)	Partially degraded RNA (RIN 2–6)
1	15 min at 94°C	7–8 min at 94°C
2	2 min at 75°C	2 min at 75°C
3	2 min at 70°C	2 min at 70°C
4	2 min at 65°C	2 min at 65°C
5	2 min at 60°C	2 min at 60°C
6	2 min at 55°C	2 min at 55°C
7	2 min at 37°C	2 min at 37°C
8	2 min at 25°C	2 min at 25°C
9	Hold at 4°C	Hold at 4°C

3. Refer to the *NEBNext Ultra II Directional RNA Library Prep Kit for Illumina Instruction Manual* and immediately proceed to “First Strand cDNA Synthesis Reaction”.

Note: “First Strand cDNA Synthesis Reaction” is chapter 4.2 in the instruction manual.

4. Follow the *NEBNext Ultra II Directional RNA Library Prep Kit for Illumina Instruction Manual* to perform all remaining library construction steps.

Revision History

Date	Changes
06/2020	Initial release



Scan QR code for handbook.

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual.

Trademarks: QIAGEN®, Sample to Insight®, QIAseq®, FastSelect™ (QIAGEN Group); Illumina® (Illumina, Inc.); NEB®, NEBNext® (New England Biolabs, Inc.). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.

06/2020 HB-2775-001 © 2020 QIAGEN, all rights reserved.