

Product Sheet

RNeasy[®] Mini QIAcube[®] Kit

For use on QIAcube Connect and classic QIAcube instruments.

RNeasy Mini QIAcube Kit (240)	Cat. no. 74116
RNeasy Mini Spin Columns	240
Rotor Adapters	240
Elution Tubes	240
Buffer RLT*	220 ml
Buffer RW1 *	220 ml
Buffer RPE† (concentrate)	65 ml
RNase-Free Water	50 ml
Product sheet	1

* Contains a guanidine salt. Not compatible with disinfectants containing bleach. See other side for safety information.

† Before using for the first time, add 4 volumes of ethanol (96–100%) as indicated on the bottle to obtain a working solution.

Storage

The RNeasy Mini QIAcube Kit should be stored dry at room temperature (15–25°C) and is stable for at least 9 months under these conditions, if not otherwise stated on the label.

Intended Use

The RNeasy Mini QIAcube Kit is intended for molecular biology applications. This product is not intended for the diagnosis, prevention, or treatment of a disease. For more information about the automated procedure, see the relevant protocol sheet available at www.qiagen.com/MyQIAcube.

Quality Control

In accordance with QIAGEN's ISO-certified Quality Management System, each lot of RNeasy Mini QIAcube Kit is tested against predetermined specifications to ensure consistent product quality.

Safety Information

When working with chemicals, always wear a suitable lab coat, disposable gloves and protective goggles. For more information, please consult the appropriate safety data sheets (SDSs). These are available online in convenient and compact PDF format at www.qiagen.com/Support/MSDS.aspx where you can find, view and print the SDS for each QIAGEN® kit and kit component.

CAUTION 	CAUTION: DO NOT add bleach or acidic solutions directly to the sample-preparation waste.
---	---

Buffer RLT contains guanidine thiocyanate and Buffer RW1 contains a small amount of guanidine thiocyanate. Guanidine salts can form highly reactive compounds when combined with bleach. If liquid containing these buffers is spilt, clean with suitable laboratory detergent and water. If the spilt liquid contains potentially infectious agents, clean the affected area first with laboratory detergent and water, and then with 1% (v/v) sodium hypochlorite.

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual. QIAGEN kit handbooks and user manuals are available at www.qiagen.com or can be requested from QIAGEN Technical Services or your local distributor.

Document Revision History

Date	Changes
May 2020	Updated text, ordering information and intended use for QIAcube Connect.

Trademarks: QIAGEN®, QIAcube®, RNeasy® (QIAGEN Group).
1120349 HB-0379-003 © 2020 QIAGEN, all rights reserved.