

QIAcuity® *E. coli* resDNA Quant Standard

QIAcuity *E. coli* resDNA Quant Standard consists of QIAcuity *E. coli* resDNA Quant Standard DNA and Rehydration Buffer. The QIAcuity *E. coli* resDNA Quant Standard DNA is a dPCR-verified absolute quantification standard that can be used in combination with the QIAcuity *E. coli* resDNA Quant Kit (Cat. No. 250220) for validation of quantitation accuracy or bridging studies. The kit is shipped on dry ice and should be stored protected from light at –30 to –15°C in a constant-temperature freezer upon receipt. Under these conditions the kit components are stable, without showing any reduction in performance and quality, until the expiry date indicated on the label.

Further information

- QIAcuity *E. coli* resDNA Quant Kit: www.qiagen.com/QIAcuity-EColi-resDNA-Quant
- *QIAcuity User Manual Extension: QIAcuity Application Guide*: www.qiagen.com/HB-2839
- *QIAcuity User Manual*: www.qiagen.com/HB-2717
- Safety Data Sheets: www.qiagen.com/safety
- Technical assistance: support.qiagen.com

Kit content

Component	Quantity	Cap color	Storage (°C)
QIAcuity <i>E. coli</i> resDNA Quant Standard (1000 pg)(lyophilized)	1	Red	–30 to –15
Rehydration Buffer, 1.5 mL	1	White	–30 to –15

Notes before starting

To maintain a working environment free of external DNA contamination, we recommend the following precautions for accurate and reproducible dPCR results:

- Wear lab coats, goggles, and gloves throughout the procedure.
- Decontaminate your dPCR workspace and labware (pipets, tube racks, etc.) before each new use to render any DNA contamination ineffective in dPCR.
- Store sample materials and control templates separately from other reagents. Physically separate dPCR setup workspaces from post-dPCR processing operations.

- Use fresh QIAcuity nanoplates and PCR-grade labware. Use sterile filter-tip pipettes.
- Do not remove the sealer foil from previously used QIAcuity nanoplates that releases dPCR product DNA into the air and contaminate results.
- Do not remove the QIAcuity nanoplate from its protective sealed bag until immediately before use.
- Pipetting accuracy and precision affect the consistency of results. Make sure that no bubbles are pipetted into the wells of the QIAcuity nanoplate. Use sterile filter-tip pipettes.
- At least one No Template Control (NTC) sample should be included in the runs to detect any external DNA contamination.
- DNA samples with ≥ 20 kb average length should be fragmented by restriction digestion before partitioning. Enzymatic fragmentation of larger DNA ensures even distribution of template throughout the QIAcuity nanoplate, which in turn leads to accurate and precise quantification.
- The following validated enzymes will not cut within the amplified sequence. It is sufficient to digest DNA templates in 10 min at room temperature (15–25°C) when added directly to the reaction mix at the indicated concentrations.

Validated restriction enzymes

6-cutter restriction enzymes

<i>EcoRI</i>	0.25 U/ μ L EcoRI-HF [®] , NEB [®]	<i>PvuII</i>	0.025 U/ μ L PvuII, NEB
	0.025 U/ μ L Anza™ 11 EcoRI, Thermo Fisher Scientific (TFS)		0.025 U/ μ L Anza 52 PvuII, TFS

Procedure

Rehydration of the reagents

Add 100 μ L Rehydration Buffer to the QIAcuity *E. coli* resDNA Quant Standard to produce a final concentration of 10 μ g/ μ L. Vortex and spin briefly after reconstitution. Afterwards, incubate for 20 min at 37°C.

Reaction setup

1. Thaw the *E. coli* resDNA Quant Standard kit and *E. coli* resDNA Quant kit components: *E. coli* resDNA Quant Master Mix (2x), internal control and dPCR qualified water. Vigorously mix the *E. coli* resDNA Quant Master Mix (2x), *E. coli* resDNA Quant Standard and the individual solutions. Centrifuge briefly to collect liquids at the bottom of the tubes. The table

below provides exemplary loading amounts and dilutions for *E. coli* resDNA Quant Standard DNA.

QIAcuity <i>E. coli</i> resDNA Quant Standard Volume* (10 pg/μL) in 40 μL reaction	Final amount in 40 μL reaction (pg)	Final concentration in 40 μL reaction (pg/μL)
18 μL	180	4.5
10 μL	100	2.5
4 μL	40	1
2 μL	20	0.5
0.4 μL	4	0.1
0.2 μL	2	0.05
0.1 μL	1	0.025

*After rehydration.

- Prepare a reaction mix for the number of reactions needed according to the table hereafter. Due to hot-start, it is not necessary to keep samples on ice during reaction setup or while programming the QIAcuity instrument. The stringency of hot-start, along with other proprietary chemical components in *E. coli* resDNA Quant Master Mix (2x) is essential for delivering highest performance in residual DNA quantification.

Component	Reagent/sample volume Nanoplate 26k (24-well)	Final concentration
QIAcuity <i>E. coli</i> resDNA Quant Master Mix (2x)	20 μL	1x
QIAcuity <i>E. coli</i> resDNA Quant Internal Control	1 μL (recommended)	100 ± 20 copies/μL *
dPCR Qualified Water	Variable	–
QIAcuity <i>E. coli</i> resDNA Quant Standard†	Variable	–
Restriction enzyme	Up to 1 μL	0.025–0.25 U/μL
Total reaction volume	40 μL	–

* Expected dPCR result when 1 μL of internal control is added to the 40 μL reaction volume. It is recommended for maximal precision to add the internal control directly in the master mix.

† QIAcuity *E. coli* resDNA Quant Standard loading amounts may vary according to the experimental setup and should not exceed 50 pg per reaction.

- Vortex gently and spin down the reaction mix.
- Dispense appropriate volumes of the reaction mix into the wells of a standard PCR plate. Then, add template DNA into each well that contains the reaction mix. Make sure all components are mixed well. Centrifuge briefly.
- Transfer the content of each well from the standard PCR plate to the wells of the nanoplate avoiding air bubbles.

6. Seal the nanoplate properly using the QIAcuity Nanoplate Seal provided in the QIAcuity Nanoplate Kits.

Note: For exact sealing procedure, please see the *QIAcuity User Manual*.

7. If a restriction enzyme for DNA digestion has been included in the reaction, leave the plate for 10 min at room temperature.

Thermal cycling and imaging conditions

1. Program the cycler of the QIAcuity instrument according to the following table:

Step	Time	Temperature (°C)
PCR initial heat activation	2 min	95
Three-step cycling (45 cycles)		
Denaturation	15 s	95
Annealing	10 s	55
Extension	30 s	60

2. Adjust the QIAcuity instrument according to the recommended Imaging settings:

Target	Detection channel	Exposure/gain
Target assay (<i>E. coli</i>)	Green	500/6
Internal Control	Yellow	500/6

3. Place the nanoplate into the QIAcuity instrument and start the dPCR program.

Note: For details, please see the *QIAcuity User Manual*.

Analysis

Use absolute quantification in QIAcuity Software Suite to calculate the target *E. coli* DNA concentration in the reaction in copies/μL. Use following table for calculating residual *E. coli* DNA amounts in fg/μL.

Kit	Target copy number	Amplicon size	Conversion factor (copies/μL to fg/μL)
QIAcuity <i>E. coli</i> resDNA Quant Kit	7	<200 bp	0.35
<i>E. coli</i> DNA concentration (copies/μL)		<i>E. coli</i> DNA concentration (fg/μL)	
	10		3.5
	20		7
	100		35
	200		70
	1000		350

Document Revision History

Date	Changes
11/2022	Initial release



Scan QR code for handbook.

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual.
Trademarks: QIAGEN®, Sample to Insight, QIAcuity® (QIAGEN Group); EcoRI-HF®, NEB® (New England Biolabs, Inc.). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.
11/2022 HB-3243-001 © 2022 QIAGEN, all rights reserved.