April 2025

Important Note

Plastic Change for EZ2® DNA Investigator® Sep&Prep Kit

Dear valued customer,

This is to notify you that, due to supplier discontinuation of the D-tube caps, modifications have been made to the kit components of EZ2 DNA Investigator Sep&Prep Kit (cat. no. 952134), resulting in a minor update to the associated protocol.

Effective from kit lot no. 184019618, D-tubes (50) (mat. no. 1136717) will be replaced by D-Tubes with caps (50) (mat. no. 1139728). Consequently, D-tube caps (50) (mat. no. 1136718) will no longer be included in the kit.

Please note that the "Pretreatment Protocol for Sexual Assault Samples" has been updated to reflect these changes. Following the initial sample lysis for 1 hour at 56°C at 900 rpm, and a brief centrifugation to remove any droplets from the cap interior, open the D-tube and carefully cut off the lid from the lower hinge. It is essential to cut the hinge as close as possible to the reaction vessel. All other steps of the protocol remain unchanged.

The corresponding *EZ2 DNA Investigator Sep&Prep Handbook* (**www.qiagen.com/HB-2432**) and *Quick Start Protocol* (**www.qiagen.com/HB-3661**) will change accordingly.

If you have any questions, you can reach us by phone or get online support. Visit **www.qiagen.com/support** for localized contact details.

Sincerely,

QIAGEN

Trademarks: QIAGEN®, Sample to Insight®, EZ2®, Investigator® (QIAGEN Group). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.
04/2026 HB-3813-01 © 2026 QIAGEN, all rights reserved.