

Labware List — QIASymphony® DSP AXpH DNA Kit




Sample and eluate tubes/racks that can be used with the QIASymphony DSP AXpH DNA Kit and the QIASymphony SP (software version 4.0; labware package SOW-516-6)

Version management

This document is the QIASymphony DSP AXpH DNA Kit Labware List for labware package SOW-516-6, Version 1, Revision 1.

Important: Before using this Labware List, confirm that it is consistent with the labware package installed on your QIASymphony system.


Legend

-  Recommended: These tubes are for use in combination with this protocol.
-  User defined: User is responsible for defining and/or validating as the user deems appropriate.
-  Not recommended: These tubes are not for use in combination with this protocol.
- x/■ Minimum sample volume (μ l) required per sample per protocol (denoted by x)/clot detection possible.
- x/□ Minimum sample volume (μ l) required per sample per protocol (denoted by x)/clot detection not possible.

Note: Be aware that other tubes are available that can be used in combination with other QIASymphony Kits, but are not recommended for any protocols that can be used with this kit.

“Sample” drawer, tube carrier

Note: Ensure that you remove swabs before using tubes on the QIASymphony SP.

					Protocol
Supplier	Material	Example cat. no.	Name in touchscreen	Insert	AXpH
BD*	14 ml Falcon polystyrene round-bottom tube 17 x 100 mm	352051	BD#352051 FalconPP 17x100	No insert needed	

* BD was the previous supplier of this tube and Corning Inc. is now the new supplier.

Labware List — QIASymphony DSP AXpH DNA Kit

“Sample” drawer, tube carrier (proteinase K)

Note: Applies only to SP2000_V2_DSP protocol.

Note: The tube carrier containing proteinase K must be placed in slot A of the “Sample” drawer of the QIASymphony SP.

					Protocol
Supplier	Material	Example cat. no.	Name in touchscreen	Insert	AXpH
BD*	14 ml Falcon polystyrene round-bottom tube 17 x 100 mm	352051	BD#352051 FalconPP 17x100	No insert needed	

* BD was the previous supplier of this tube and Corning Inc. is now the new supplier.

Labware List — QIASymphony DSP AXpH DNA Kit

“Eluate” drawer

						Protocol
Supplier	Material	Example cat. no.	Category	Name in touchscreen	Adapter on Elution slots [†]	AXpH
Greiner Bio-One	Polystyrene Microplate, 96 well, u-bottom	650161	Micro Plate	GR#650161 *MTP96 RB	96-Well Round Bottom QS (cooling slot 1)	Green
				GR#650161 MTP96 RB	No adapter needed (non-cooling slots 2-3)	Yellow
QIAGEN	Elution Microtubes CL 96	19588	Deep Well	QIA#19588 *EMTR	Elution Microtube Rack QS (cooling slot 1)	Green
				QIA#19588 EMTR	No adapter needed (non-cooling slots 2-3)	Yellow

* Indicates labware that can be cooled using a cooling adapter with bar code (transferable and usable on QIASymphony AS).

** Indicates labware that can be cooled using a cooling adapter without bar code (non-transferable and not usable on QIASymphony AS).

† Do not use 96-well plates on “Elution slot 4” because the robotic arm cannot access all positions.

For ordering information, visit www.qiagen.com/goto/QIASymphony.

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual. QIAGEN kit handbooks and user manuals are available at www.qiagen.com or can be requested from QIAGEN Technical Services or your local distributor.

Trademarks: QIAGEN®, QIASymphony® (QIAGEN Group); BD™ (Becton, Dickinson and Company); Corning®, Falcon® (Corning, Inc.); Bio-One® (Greiner Bio-One GmbH). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.

Jan-16 HB-0300-L05-001 © 2012–2016 QIAGEN, all rights reserved.

