

March 2021

QuantiFERON[®] SARS-CoV-2 Starter Set Blood Collection Tubes

Instructions for Use (Handbook)



For Research Use Only

Not for use in diagnostic procedures

REF

626115



www.qiagen.com



QIAGEN, 19300 Germantown Road, Germantown, MD 20874, USA
Phone: +1-800-426-8157

R2 **MAT**

1122890

Contents

Intended Use	4
Materials Provided	5
Kit contents	5
Warnings and Precautions	6
Procedures	7
Stage 1: Blood collection and hold time options	7
Stage 2: Post-incubation of BCTs and harvesting of plasma	13
Symbols	14
Troubleshooting Guide	15
Contact Information	17
Ordering Information	18
Document Revision History	19

Intended Use

The QuantiFERON SARS-CoV-2 Starter Set Blood Collection Tubes (QFN SARS-CoV-2 BCTs) consists of two Antigen tubes, SARS-CoV-2 Ag1 and SARS-CoV-2 Ag2, that use a combination of antigens specific to SARS-CoV-2 to stimulate lymphocytes in heparinized whole blood involved in cell-mediated immunity. Plasma from the stimulated samples can be used for detection of IFN- γ . Detection of IFN- γ can be done using QuantiFERON ELISA.

The QFN SARS-CoV-2 BCTs are for Research Use Only and not for diagnostic use.

Materials Provided

Kit contents

Blood Collection Tubes		200 tubes
Catalog no.		626115
QuantiFERON SARS-CoV-2 Ag1 Tube (black cap, red ring)	Ag1	100 tubes
QuantiFERON SARS-CoV-2 Ag2 Tube (black cap, ochre ring)	Ag2	100 tubes
Language Sheet	–	1

QFN SARS-CoV-2 BCTs are designed to draw the required volume of blood for stimulation. The contents of the BCTs have been dried onto the inner walls, and it is essential that the BCTs be thoroughly mixed with the blood to resolubilize them. Blood collected directly into the QFN SARS-CoV-2 BCTs must be transferred to a 37°C incubator as soon as possible and within 16 hours of blood collection (see Direct draw into QFN SARS-CoV-2 BCTs)

Alternatively, blood may be collected into a single lithium-heparin or sodium-heparin tube for storage prior to transfer to QFN SARS-CoV-2 BCTs and incubation. Blood specimens collected in heparin tubes can be stored at room temperature (17–25°C) but held for no more than 16 hours from the time of collection prior to transfer to QFN SARS-CoV-2 BCTs and subsequent incubation (see Blood collection into a heparin tube and then transfer to QFN SARS-CoV-2 BCTs with room temperature storage and handling). Blood specimens in heparin tubes may also be stored at 2–8°C for up to 48 hours prior to transfer to the QFN SARS-CoV-2 BCTs (see Blood collection into a heparin tube and then transfer to QFN SARS-CoV-2 BCTs with refrigerated storage and handling).

Warnings and Precautions

For Research Use Only. Not for use in diagnostic procedures.

When working with chemicals, always wear a suitable lab coat, disposable gloves, and protective goggles. For more information, please consult the appropriate safety data sheets (SDSs). These are available online in PDF format at www.qiagen.com where you can find, view, and print the SDS for each QIAGEN kit and kit component.

<p>CAUTION</p> 	<p>Handle human blood as if potentially infectious. (C1)</p> <p>Observe relevant blood handling guidelines. Dispose of samples and materials in contact with blood or blood products in accordance with federal, state, and local regulations.</p>
---	--

Procedures

Stage 1: Blood collection and hold time options

See Blood Collection Options below (Figures 1–3).

Direct draw into QFN SARS-CoV-2 BCTs

1. Label BCTs appropriately. Ensure that each BCT (Ag1 and Ag2) is identifiable by its label or other means once the cap is removed.

Note: It is recommended to record the time and date of blood collection.

Important: QFN SARS-CoV-2 BCTs should be at room temperature 17–25°C (62.6–77°F) at the time of blood collection.

2. For each patient, collect 1 ml of blood by venipuncture directly into each of the QFN SARS-CoV-2 BCTs.

Important: This procedure must be performed by a trained phlebotomist.

- As 1 ml BCTs draw blood relatively slowly, keep the BCT on the needle for 2–3 seconds once the BCT appears to have completed filling. This will ensure that the correct volume is drawn.
- The black mark on the side of the BCTs indicates the validated range of 0.8 to 1.2 ml. If the level of blood in any BCT is outside of the indicator mark, a new blood sample should be obtained. Under- or over-filling of the BCTs outside of the 0.8 to 1.2 ml range may lead to erroneous results.
- If a “butterfly needle” is being used to collect blood, a “purge” tube should be used to ensure that the tubing is filled with blood prior to the QFN SARS-CoV-2 BCTs being used.
- QFN SARS-CoV-2 BCTs can be used up to an altitude of 2650 feet (810 meters) above sea level.
- If using QFN SARS-CoV-2 BCTs outside altitude ranges or if low blood draw volume occurs, users can collect blood with a syringe, and immediately transfer 1 ml to each

of the BCTs. For safety reasons, this is best performed by removing the syringe needle, ensuring appropriate safety procedures, removing the caps from the QFN SARS-CoV-2 BCTs and adding 1 ml of blood to each (to the black mark on the side of the BCT label which indicates the validated range of 0.8 to 1.2 ml). Replace the caps securely and mix as described below. Ensure each BCT (Ag1 and Ag2) is identifiable by its label or other means once the cap is removed.

3. Immediately after filling the BCTs, shake them ten (10) times just firmly enough to ensure that the entire surface of the BCT is coated with blood. This will dissolve antigens on the BCTs walls.

Important: Over vigorous shaking may cause gel disruption and could lead to aberrant results.

4. Following labelling, filling, and shaking, the BCTs must be transferred to a $37^{\circ}\text{C} \pm 1^{\circ}\text{C}$ incubator as soon as possible, and within 16 hours of collection. Prior to incubation, maintain BCTs at room temperature ($22^{\circ}\text{C} \pm 5^{\circ}\text{C}$ [$71.6^{\circ}\text{F} \pm 9^{\circ}\text{F}$]). If QFN SARS-CoV-2 BCTs are not incubated at 37°C directly after blood collection and shaking, invert the BCTs to mix 10 times prior to incubation at 37°C .
5. Incubate the QFN SARS-CoV-2 BCTs upright at $37^{\circ}\text{C} \pm 1^{\circ}\text{C}$ for 16 to 24 hours.

Note: The incubator does not require CO_2 or humidification.




Figure 1. Blood collection option: Direct draw into QFN SARS-CoV-2 BCTs and hold at room temperature. The total time from blood draw in QFN SARS-CoV-2 BCTs to 37°C incubation must not exceed 16 hours.

Blood collection into a heparin tube and then transfer to QFN SARS-CoV-2 BCTs with room temperature storage and handling

1. Blood may be collected in a blood collection tube containing heparin as the anticoagulant and then transferred to QFN SARS-CoV-2 BCTs. Only use heparin as the blood containing coagulant because other anticoagulants interfere with the assay. Label tubes appropriately.

Note: It is recommended to label the tube with the time and date of the blood collection.

2. Fill a heparin blood collection tube (≥ 3 ml for the QFN SARS-CoV-2 BCTs, and additional for any other tubes being tested) and gently mix by inverting the BCT several times to dissolve the heparin.

Important: This procedure must be performed by a trained phlebotomist.

3. Blood collected in heparin tube must be maintained at room temperature (17–25°C [62.6–77°F]) for no more than 16 hours from the time of collection prior to transfer to QFN SARS-CoV-2 BCTs and subsequent incubation.
4. Transfer of blood specimen from a heparin tube to QFN SARS-CoV-2 BCTs.

- Label each QFN SARS-CoV-2 BCTs appropriately.

Note: Ensure each BCT (Ag1 and Ag2) is identifiable by its label or other means once the cap is removed. It is recommended to transfer the recorded time and date of blood collection from the heparin tubes to the QFN SARS-CoV-2 BCTs.

- Samples must be evenly mixed by gentle inversion before dispensing into QFN SARS-CoV-2 BCTs.
 - Dispensing should be performed aseptically, ensuring appropriate safety procedures, removing the caps from the 2 QFN SARS-CoV-2 BCTs, and adding 1 ml of blood to each BCT. Replace the BCT caps securely and mix as described in the next steps.
5. Mix BCTs. Immediately after filling the QFN SARS-CoV-2 BCTs, shake them ten (10) times just firmly enough to ensure that the entire inner surface of the BCT is coated with blood. This will dissolve antigens on BCT walls.

Important: Overly vigorous shaking may cause gel disruption and could lead to aberrant results.

6. Incubate the QFN SARS-CoV-2 BCTs upright at $37^{\circ}\text{C} \pm 1^{\circ}\text{C}$ for 16 to 24 hours.

Note: The incubator does not require CO_2 or humidification.




Figure 2. Blood collection option: Draw into heparin tube and hold at room temperature. The total time from blood draw in heparin tube to 37°C incubation must not exceed 16 hours.

Blood collection into a heparin tube and then transfer to QFN SARS-CoV-2 BCTs with refrigerated storage and handling

1. Fill a heparin blood collection tube (≥ 3 ml for the QFN SARS-CoV-2 BCTs, and additional for any other tubes being tested) and gently mix by inverting the tube several times to dissolve the heparin.

Important: This procedure should be performed by a trained phlebotomist.

2. Before refrigeration, blood drawn into heparin tube may be held at room temperature (17–25°C) up to 3 hours after blood collection.
3. Blood drawn into heparin tube may be refrigerated (2–8°C) up to 48 hours.
4. After refrigeration, heparin tube must equilibrate to room temperature (17–25°C) for 1 hour prior to transfer to QFN SARS-CoV-2 BCTs.
 - Aliquoted QFN SARS-CoV-2 BCTs should be placed in the 37°C incubator within 2 hours of removing the heparin tube from 2–8°C.
 - Label each QFN SARS-CoV-2 BCT appropriately.

Note: Ensure each BCT (Ag1 and Ag2) is identifiable by its label or other means once the cap is removed. It is recommended to transfer the recorded time and date of blood collection from the heparin tubes to the QFN SARS-CoV-2 BCTs.
 - Samples must be evenly mixed by gentle inversion before dispensing into QFN SARS-CoV-2 BCTs.
 - Dispensing should be performed aseptically, ensuring appropriate safety procedures, removing the caps from the 2 QFN SARS-CoV-2 BCTs, and adding 1 ml of blood to each BCT. Replace the BCT caps securely and mix as described below.
5. Following labelling, filling, and shaking, the BCTs must be transferred to a 37°C \pm 1°C incubator within 2 hours of removing heparin tubes from 2–8°C. If QFN SARS-CoV-2 BCTs are not incubated at 37°C directly after blood collection and shaking, invert the BCTs to mix 10 times prior to incubation at 37°C. (See Figures 1–3 for blood collection options.)




Figure 3. Blood collection option: Draw into heparin tube and hold at 2–8°C. The total time from blood drawing heparin tube to 37°C incubation must not exceed 53 hours. Note: Aliquoted QFN SARS-CoV-2 BCTs should be placed in a 37°C incubator within 2 hours of removing heparin tube from 2-8°C.

Stage 2: Post-incubation of BCTs and harvesting of plasma

Things to do before starting

- Prior to harvesting plasma, samples in QFN SARS-CoV-2 BCTs must be incubated at 37°C for 16–24 hours. The incubator does not require CO₂ or humidification.

Procedure

1. After incubation at 37°C ± 1°C, the BCTs may be held between 4°C and 27°C for up to 3 days prior to centrifugation.
2. After incubation of the BCTs at 37°C ± 1°C, harvesting of the plasma is facilitated by centrifuging the BCTs for 15 minutes at 2000 to 3000 RCF (g). The gel plug will separate the cells from the plasma. If this does not occur, the BCTs should be re-centrifuged.
3. It is possible to harvest the plasma without centrifugation, but additional care is required to remove the plasma without disturbing the cells.
4. Plasma samples should only be harvested using a pipet.

Important: After centrifugation, avoid pipetting plasma up and down or mixing plasma by any means prior to harvesting. At all times, take care not to disturb material on the surface of the gel.

Plasma samples can be stored in centrifuged QFN SARS-CoV-2 BCTs for up to 28 days at 2–8°C, or harvested plasma samples can be stored for up to 28 days at 2–8°C. Harvested plasma samples can also be stored below –20°C (preferably less than –70°C) for extended periods.

Symbols

The following symbols may appear on the packaging and labelling:

Symbol	Symbol definition
	Legal manufacturer
	Batch code
	Catalog number
	Global Trade Item Number
	Material number (i.e., component labeling)
	Use by
	Temperature limitation
	Consult instructions for use
	Do not reuse
	Sterilized using irradiation
	For Research Use Only. Not for use in diagnostic procedures.
	Caution

Troubleshooting Guide

This troubleshooting guide may be helpful in solving any problems that may arise. For technical assistance and more information, please see our Technical Support Center at www.qiagen.com/support (for contact information, visit www.qiagen.com).

Comments and suggestions

Underfilling of BCT

- | | |
|---|---|
| a) BCT removed from the needle too soon. | As 1 ml BCTs draw blood relatively slowly, keep the BCT on the needle for 2–3 seconds once the BCT appears to have completed filling. This will ensure that the correct volume is drawn. |
| b) Blood drawn outside the recommended altitude of 2650 feet (810 meters) above sea level | QFN SARS-CoV-2 BCTs can be used up to an altitude of 2650 feet (810 meters) above sea level.
If using QFN SARS-CoV-2 BCTs outside altitude ranges or if low blood draw volume occurs, users can collect blood with a syringe, and immediately transfer 1 ml to each of the BCTs. |
| c) Tubing not primed while using butterfly needle | If a “butterfly needle” is used to collect blood, a “purge” tube should be used to ensure that the tubing is filled with blood prior to the QFN BCTs being used. |
| d) BCTs are past their expiration date | BCTs must be used within the expiration date printed on the tube label. |

Overfilling of BCT

- | | |
|--|---|
| Tube not at room temperature during blood collection | BCTs should be at room temperature 17–25°C (62.6–77°F) at the time of blood collection. |
|--|---|

Blood clots

- | | |
|---------------------|--|
| Insufficient mixing | Immediately after filling the BCTs, shake them ten (10) times just firmly enough to make sure the entire inner surface of the BCT is coated with blood. This will dissolve antigens on the BCTs walls. |
|---------------------|--|

Plasma not separated by gel

- | | |
|---|--|
| Insufficient centrifugation speed or time | Harvesting of the plasma is facilitated by centrifuging the BCTs for 15 minutes at 2000–3000 RCF (g). The gel plug will separate the cells from the plasma. If this does not occur, the BCTs should be re-centrifuged. |
|---|--|

Comments and suggestions

Gel disruption

Tubes shaken too vigorously

Immediately after filling the BCTs, shake them ten (10) times just firmly enough to make sure the entire inner surface of the BCT is coated with blood. This will dissolve antigens on the BCTs walls.

Important: Over vigorous shaking may cause gel disruption and could lead to aberrant results.

Contact Information

For technical assistance and more information, please call toll-free 800-362-7737, see our Technical Support Center at www.qiagen.com/contact or contact one of the QIAGEN Technical Service Departments (see back cover or visit www.qiagen.com).

Ordering Information

Product	Contents	Cat. no.
QuantiFERON SARS-CoV-2 Starter Pack	Contains QuantiFERON Starter Set (cat. no. 626115 SARS-CoV-2 Ag1 tube and SARS-CoV-2 Ag2 tube) and QuantiFERON Control Set (cat. no. 626015 Nil and Mitogen)	626715
Related Products		
QuantiFERON Control Set	Contains Nil tube and Mitogen tube	626015
QuantiFERON Monitor Direct	Contains Monitor Direct tube	626315
QuantiFERON ELISA	Contains Microtiter Plate, Conjugate (100x), IFN Gamma Standard, Green Diluent, Wash Buffer (x20), Enzyme Substrate Solution, and Enzyme Stopping Solution	626410

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual. QIAGEN kit handbooks and user manuals are available at www.qiagen.com or can be requested from QIAGEN Technical Services or your local distributor.

Document Revision History

Date	Changes
R1, November 2020	Initial release
R2, January 2021	Updated cover page to add symbols for Temperature limitation and Do not reuse Updated Symbols section to add Material number Updated Ordering Information section

This page intentionally left blank.

This page intentionally left blank.

This page intentionally left blank.

Limited License Agreement for QuantiFERON® SARS-CoV-2 Starter Set Blood Collection Tubes

Use of this product signifies the agreement of any purchaser or user of the product to the following terms:

1. The product may be used solely in accordance with the protocols provided with the product and this handbook and for use with components contained in the kit only. QIAGEN grants no license under any of its intellectual property to use or incorporate the enclosed components of this kit with any components not included within this kit except as described in the protocols provided with the product, this handbook, and additional protocols available at www.qiagen.com. Some of these additional protocols have been provided by QIAGEN users for QIAGEN users. These protocols have not been thoroughly tested or optimized by QIAGEN. QIAGEN neither guarantees them nor warrants that they do not infringe the rights of third-parties.
2. Other than expressly stated licenses, QIAGEN makes no warranty that this kit and/or its use(s) do not infringe the rights of third-parties.
3. This kit and its components are licensed for one-time use and may not be reused, refurbished, or resold.
4. QIAGEN specifically disclaims any other licenses, expressed or implied other than those expressly stated.
5. The purchaser and user of the kit agree not to take or permit anyone else to take any steps that could lead to or facilitate any acts prohibited above. QIAGEN may enforce the prohibitions of this Limited License Agreement in any Court, and shall recover all its investigative and Court costs, including attorney fees, in any action to enforce this Limited License Agreement or any of its intellectual property rights relating to the kit and/or its components.

For updated license terms, see www.qiagen.com.

Trademarks: QIAGEN®, Sample to Insight®, QFN®, QuantiFERON® (QIAGEN Group). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.

For up-to-date licensing and product-specific disclaimers, see the *QuantiFERON SARS-CoV-2 Set BCT Package Insert*.

1122890 01-2021 © 2021 QIAGEN, all rights reserved.

