

June 2020

Labware List — QIASymphony[®] PAXgene[®] Blood ccfDNA Kit

Sample and eluate tubes/racks that can be used with the QIASymphony PAXgene Blood ccfDNA Kit and the QIASymphony SP (software version 5.0; labware package SOW-516-8)

This document is the QIASymphony PAXgene Blood ccfDNA Kit *Labware List* R1 for labware package SOW-516-8.

General information

For Research Use Only. Not for use in diagnostics procedures. No claim or representation is intended to provide information for the diagnosis, prevention, or treatment of a disease.

Important: Before using this Labware List, confirm that it is consistent with the labware package installed on your QIAasymphony system.

Legend

-  Recommended: These tubes are for use in combination with this protocol.
-  User-defined: User is responsible for defining and/or validating as the user deems appropriate.
-  Not recommended: These tubes are not for use in combination with this protocol.

- x/■ Minimum sample volume (µl) required per sample per protocol (denoted by x)/clot detection possible.
- x/□ Minimum sample volume (µl) required per sample per protocol (denoted by x)/clot detection not possible.

Note: Be aware that other tubes are available that can be used in combination with other QIAasymphony Kits, but are not recommended for any protocols that can be used with this kit.

“Sample” drawer, tube carrier

Supplier	Material	Example cat. no.	Name in touchscreen	Insert	Protocol			
					PAXcircDNA STA 2400	PAXcircDNA LAF 2400	PAXcircDNA STA 4800*	PAXcircDNA LAF 4800*
BD [†]	14 ml Falcon® polystyrene round-bottom tube 17 x 100 mm	352051	BD#352051 FalconPP 17x100	No insert needed	2800 [‡] /■		5300 [‡] /■	
BD [†]	14 ml Falcon polystyrene round-bottom tube 17 x 100 mm	352051	BD_FIX_#352051 FalconPP 17x100 [§]	No insert needed	2500 [§] /□		4900 [§] /□	
BD [†]	10 ml Falcon polystyrene round-bottom tube 16 x 100 mm	768115	BD_#768115 PAXgene ccfDNA 16x100 [§]	No insert needed	■		■	

* Make sure that the correct sample volume is used. A sample volume (including void volume) that is less than required is recognized only at the second sample transfer.

[†] BD was the previous supplier of this tube; Corning® Inc. is now the new supplier.

[‡] Using less sample volume than listed results in “unclear” flagging of samples (error code 140043 “EnableLessSample mode”). The “Enable less sample” mode is designed to use all available liquid in combination with liquid-level detection and clot detection when less volume than listed above is detected during sample transfer.

[§] The tube is verified, but its use will lead to an unclear flagging of samples when the configuration parameter “Check combination of protocol and recommended labware during run definition?” is active.

[§] Reduced minimum sample volume (µl) to minimize dead volumes. FIX labware is designed for this purpose and does not support liquid-level detection or clot detection. FIX sample tubes impose aspiration restrictions; the sample is aspirated at a defined height in the tube, which is defined by the volume of sample to be transferred. Therefore, make sure to use the volume listed above.

Note: It is possible to process tubes with and without liquid-level detection within the same batch/run.

“Sample” drawer, tube carrier (ProteinaseK) slot-A

Note: Preparation of ProteinaseK volume is described in the corresponding protocol sheet.

Note: The tube carrier containing the ProteinaseK must be placed in slot-A of the “Sample” drawer.

Supplier	Material	Example cat. no.	Name in touchscreen	Insert	Protocol			
					PAXcircDNA STA 2400	PAXcircDNA LAF 2400	PAXcircDNA STA 4800*	PAXcircDNA LAF 4800*
BD*	14 ml Falcon polystyrene round-bottom tube 17 x 100 mm	352051	BD#352051 FalconPP 17x100	No insert needed				

* BD was the previous supplier of this tube; Corning Inc. is now the new supplier.

“Eluate” drawer

Supplier	Material	Example cat. no.	Category	Name in touchscreen	Adapter on Elution slots	Protocol			
						PAXcircDNA STA 2400	PAXcircDNA LAF 2400	PAXcircDNA STA 4800*	PAXcircDNA LAF 4800*
QIAGEN	Elution Microtubes CL 96	19588	Deep Well	QIA#19588 *EMTR	Elution Microtube Rack QS, 96-wells, cat-no.: 9020730 (cooling slot 1)				
Eppendorf	1.5 ml Eppendorf® LoBind Snap Cap Safe-Lock tube	0030108.051	Tube 1.5 ml	EP#0030108.051 *T1.5 Snap Cap	Micro Tube Snap Cap, 24-wells, cat-no.: 9020731 (cooling slot 1)†				
Sarstedt	1.5 ml Micro tube, PP, NON-SKIRTED	72.607	Tube, 1.5 ml Adapter V2	SAR#72.607 *T1.5 Screw	Micro Tube Screw Cap QS, 24-wells, cat-no.: 9020674 (cooling slot 1)				
Sarstedt	2.0 ml Micro tube, PP, NON-SKIRTED	72.693	Tube 2.0 ml / Tube_2.0ml AdapterV1 (no BC)	SAR#72.693 *T2.0 Screw	Micro Tube Screw Cap QS, 24-wells, cat-no.: 9020674 (cooling slot 1)				
				SAR#72.693 **T2.0 Screw	Micro Tube Screw Cap QS (cooling slot 1)				
				SAR#72.693 T2.0 Screw	24-Tube 1.5/2.0 ml QS (non-cooling slots 2–4)				
Starlab	1.5 ml Micro tube, graduated conical tube, non-skirted	E1415-2231	Tube, 1.5 ml / Tube_1.5ml AdapterV1 (no BC)	SL#E1415-2231 *T1.5 Screw	Micro Tube Screw Cap QS, 24-wells, cat-no.: 9020674 (cooling slot 1)				
				SL#E1415-2231 **T1.5 Screw	Micro Tube Screw Cap QS (cooling slot 1)				
				SL#E1415-2231 T1.5 Screw	1.5/2.0 ml QS (non-cooling slots 2–4)				

* Indicates labware that can be cooled using a cooling adapter with bar code.

** Indicates labware that can be cooled using a cooling adapter without bar code (non-transferable and not usable on QIA Symphony AS).

† Labware that cannot be transferred to or used on QIA Symphony AS.

For ordering information, visit www.qiagen.com/goto/QIASymphony.

Revision history

Document Revision History	
R1 06/2020	Release of new labware package

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN® kit handbook or user manual. QIAGEN kit handbooks and user manuals are available at www.qiagen.com or can be requested from QIAGEN Technical Services or your local distributor.

Trademarks: QIAGEN®, Sample to Insight®, QIASymphony® (QIAGEN Group); BD™ (Becton, Dickinson and Company); Corning®, Falcon® (Corning, Inc.); Eppendorf® (Eppendorf AG); PAXgene® (PreAnalytiX GmbH); Sarstedt®, (Sarstedt AG and Co.); Starlab® (Starlab Group). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.
06/2020 HB-2101-L03-001 © 2020 QIAGEN, all rights reserved.

Ordering www.qiagen.com/shop | Technical Support support.qiagen.com | Website www.qiagen.com