Important Note

therascreen® EGFR RGQ PCR Kit (24), REF 870121

Dear, valued therascreen EGFR RGQ PCR Kit customer.

You have been informed previously that QIAGEN has revised the Instructions for Use for *therascreen* EGFR PCR Kits to reduce any risk resulting from fluorescence artefacts leading to invalid runs or false mutation positive results, and to emphasize the importance of thorough mixing of reagents. A lower cut-off for the delta C_T (ΔC_T) and an additional manual check for the delta CT values (ΔCT) was introduced to avoid any false positive results due to such fluorescence artefacts.

In this Important Note, we would like to inform you that the automatic calling of results in the Rotor-Gene Software has been updated and new revision of the corresponding kit handbook has been released:

Revision 10 for REF 870121 therascreen® EGFR RGQ PCR Kit (24)

These new handbook revisions include all corresponding Instructions for Use changes. Please refer to the Document Revision History of the applicable *therascreen* EGFR RGQ PCR Kit Handbook for more details.

The Assay Package for automatic result calling is available for download in the *therascreen* EGFR RGQ PCR Kit product webpage at **www.qiagen.com**. Go to **Product Resources** > **Supplementary Protocols** to download the assay package.



Translations of this Important Note

QIAGEN provides the printed Important Note in the English language. Translations of the Important Note are electronically available in a portable document format (PDF). You may access the translated documents in the Product Resources tab of the corresponding product page in **www.qiagen.com**.

If you have any questions or concerns, please contact your local QIAGEN Technical Service Department. For contact information, visit the following webpages:

QIAGEN Subsidiaries: https://www.qiagen.com/about-us/contact/global-contacts/subsidiaries/

QIAGEN Commercial Partners and Importers:

https://www.qiagen.com/about-us/contact/global-contacts/distributors-and-importers/

With kind regards,

QIAGEN www.qiagen.com

Trademarks: QIAGEN[®], Sample to Insight[®], therascreen[®] (QIAGEN Group). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law. 1122316 09:2020 HB:2822-2001 © 2020 QIAGEN, all rights reserved.